

**To:** Moore, Kendall[moore.kendall@epa.gov]; Mullin, Michelle[Mullin.Michelle@epa.gov]  
**Cc:** Crawford, Jennifer[Crawford.Jennifer@epa.gov]; Pace, Christopher[Pace.Christopher@epa.gov]; Norton, Karen[Norton.Karen@epa.gov]; Wood, Kim[Wood.Kim@epa.gov]  
**From:** Dodo, Gerald  
**Sent:** Fri 4/1/2016 7:04:39 PM  
**Subject:** PCB Reporting Limit for Sky Valley PCB Inspection, OCE-009A

The wipe samples are proving to be challenging due to chemical interferences that are negatively affecting the GC system and high PCB levels that exceed the calibration range. As a result, the sample extracts will need to be analyzed at a higher reporting limit (RL), 2.5 ug instead of 0.25 ug/wipe. The RLs will be higher for samples with high level of PCBs detected that require further extract dilutions.

As discussed with Michelle, preliminary results would be helpful so we will try to have a summary for you sometime nextweek. Let me know if you have any questions.

Gerald Dodo

Supervisory Chemist

DWCO

USEPA Region 10 Laboratory

7411 Beach Dr. East

Port Orchard, WA 98366

Phone: (360) 871-8728

FAX: 360-871-8748